



This lot of product meets the accepted performance criteria recommended in the ISO 11138-1.

Product Name: Biological Indicator For use in Monitoring: Vaporized Hydrogen Peroxide (VH<sub>2</sub>O<sub>2</sub>)

Organism: *Geobacillus stearothermophilus* Derived from: Cell Line 12980

<b>LOT</b> S1-1	<b>REF</b> WGH-06	Quantity: 100 Wires
2020-02-20	2021-08-21 (YYYY-MM-DD)	

Performance Characteristics		
Population:	2.0x 10 <sup>6</sup> per Wire	
D value	Hydrogen Peroxide 55°C, 2.3 mg/L	2.0 minutes
The D value(s) were determined per the fraction negative method and are reproducible only when exposed and cultured under the exact conditions used to obtain results reported above. The user would not necessarily obtain the same results, therefore, should determine the suitability for their particular use.		

**Recovery**

After exposure to the sterilization process, aseptically transfer each carrier to an individual tube containing 5-15 mL of Soybean Casein Digest Broth (SCDB). Incubate at 55°-65°C for a minimum of 7 days or per validated reduced incubation time.

Examine tubes for growth, daily or as frequently as possible. Tubes showing typical growth of *Geobacillus stearothermophilus* will display a cream-colored sediment.

Purity: Shall not contain any contamination that would adversely affect the performance or the stability characteristics of the biological indicator.

Storage and Shelf Life			
+15°C to +30°C	Room Temperature (15° - 30°C)		Keep away from sunlight
20% to 80%	20% to 80% Relative Humidity		Protect from heat, radioactive sources, and sterilizing agents
Shelf Life	18 Months from the date of manufacture		

Disposal: Autoclave, steam at 121°C for not less than 30 minutes, or incinerate (standard microbial waste; non-pathogenic species).

\_\_\_\_\_  
QUALITY ASSURANCE APPROVAL

\_\_\_\_\_  
DATE