

Product Name: Hydrogen Peroxide Sterilization Process Indicator Ink


For use in Monitoring: Hydrogen Peroxide


Appearance: Magenta Liquid

LOT 210101-10P

REF IPRB-02






Origin of Manufacture: United States

 01-JAN-2021

 01-MAY-2021

Characteristic	Specification	Result
pH (at 25°C)	9.0 - 10.5	9.5
Viscosity (Zahn Cup #2)	15 – 25 seconds	20 seconds
Color Transition (6 minutes at 50°C ± 2°C in VH ₂ O ₂)	Red/Pink to Blue	Pass
Date of Product Performance Verification:	January 2021	

Color Transition: Color transition will vary depending on exposure conditions, application of ink in terms of thickness and product build.

Storage and Shelf Life			
	Room Temperature (15° - 30°C)		Keep away from sunlight
	20% to 80% Relative Humidity		Protect from heat, radioactive sources, and sterilizing agents
	Do not freeze.	Shelf Life	4 Months from the date of manufacture

Disposal: Refer to the Safety Data Sheet.

True Indicating will not be liable for any loss, costs or damage and shall not be liable for any incidental or consequential costs or damages.

October 18, 2021

Quality Approval

Date