



This lot of product meets the accepted performance criteria recommended as a Process Indicator for Ultra-Violet (UV).


Product Name: Indicator for Monitoring UV Disinfection Processes


For use in Monitoring: UV at a minimum of 50 mJ/cm<sup>2</sup>

Appearance: Yellow to Red, 19 mm x 64 mm Strip-Label





**LOT** R190909

**REF** UV-CS

 09-SEP-2019

 09-SEP-2022

Characteristic	Specification	Result
UV: 254 nm, 2 minutes	Yellow to Red Color Transition	Pass
Date of Product Performance Verification:	September 2019	

Storage and Shelf Life			
	Room Temperature (15°– 30°C)		Keep away from sunlight
	20% to 80% Relative Humidity		Protect from heat, radioactive sources, and sterilizing agents
Shelf Life	36 Months from the date of manufacture		

Disposal: Discard as general waste.

\_\_\_\_\_  
Quality Approval

\_\_\_\_\_  
Date