

Certificate of Analysis

This lot of product meets the accepted performance criteria recommended in ISO 11140-1: 2014 as a Type 1 Process Indicator.					
Product Name: Chemical Process Indicator for Monitoring Irradiation Processes					
LOT R180813		REF CRYR-4		-AUG-2021	
13-AUG-2018	13-AUG-2018 F		or use in Monitoring: Irradiation (Gamma)		
Appearance: Yellow to Red, 25 mm Label with Indicating Ribbon					
Performance Characteristics					
Characteristic		Specification		Result	
Radiation Cobalt ₆₀ , 10 kGy		Yellow to Red Color Transition		Pass	
Date of Product Performance Verification		August 2018			
Storage and Shelf Life					
+15°C-	Room Temperature (:	15°– 30°C)	**	Keep away from sunlight	
80%	20% to 80% Relative Humidity		**	Protect from heat, radioactive sources, and sterilizing agents	
Shelf Life	36 Months from the date of manufacture				
Disposal: Discard as general waste.					
Quality Approval		Date			

True Indicating, LLC 946 Kane Street, Suite A Toledo, Ohio 43612 www.trueindicating.com info@trueindicating.com