



This lot of product meets the accepted performance criteria recommended as a Process Indicator for Ultra-Violet (UV).


Product Name: Indicator for Monitoring UV Disinfection Processes


For use in Monitoring: UV at a minimum of 50 mJ/cm²

Appearance: Yellow to Red, 12 mm Indicator Dot





LOT R190901

REF CVYR-1

 01-SEP-2019

 01-SEP-2022

Characteristic	Specification	Result
UV: 254 nm, 2 minutes	Yellow to Red Color Transition	Pass
Date of Product Performance Verification:	October 2019	

Storage and Shelf Life			
	Room Temperature (15°– 30°C)		Keep away from sunlight
	20% to 80% Relative Humidity		Protect from heat, radioactive sources, and sterilizing agents
Shelf Life	36 Months from the date of manufacture		

Disposal: Discard as general waste.

Quality Approval

Date