



Product Name: Hydrogen Peroxide Sterilization Process Indicator Ink

For use in Monitoring: Hydrogen Peroxide

Appearance: Orange Liquid

LOT 210101-10P

REF IPRY-01

Origin of Manufacture: United States

01-JAN-2021

01-JUL-2021

Characteristic	Specification	Result
pH (at 25°C)	9.0 - 10.5	9.5
Viscosity, Zahn Cup No. 2	15-25 seconds	25 seconds
Color Transition (≥6 minutes at 50°C ± 2°C in VH ₂ O ₂)	Red to Yellow	Pass
Date of Product Performance Verification:	January 2021	

Color Transition: Color transition will vary depending on exposure conditions, application of ink in terms of thickness and product build.

Storage and Shelf Life			
	Room Temperature (15°-30°C)		Keep away from sunlight
	20% to 80% Relative Humidity		Protect from heat, radioactive sources, and sterilizing agents
	Do not freeze.	Shelf Life	6 Months

Disposal: Refer to the Safety Data Sheet.

True Indicating will not be liable for any loss, costs or damage and shall not be liable for any incidental or consequential costs or damages.

October 18, 2021

Quality Approval

Date